



Zettachring 6, 70567 Stuttgart, Germany  
 Tel: +49-711-13257-0 Fax: +49-711-13257-89  
 E-Mail: info.tbq@tosoh.com  
 Web: www.tosohbioscience.com

3604 Horizon Drive, Suite 100, King of Prussia, PA 19406, USA  
 Tel: +1 800-366-4875 Fax: +1 610-272-3028  
 E-Mail: info.tbl@tosoh.com  
 Web: www.tosohbioscience.com

## OPERATING CONDITIONS and SPECIFICATIONS

### TSK-GEL® PW<sub>XL</sub>-CP Products

<b>Part Numbers:</b>	21873	7.8 mm ID X 30.0 cm L	G3000 PW <sub>XL</sub> -CP	7 µm
	21874	7.8 mm ID X 30.0 cm L	G5000 PW <sub>XL</sub> -CP	10 µm
	21875	7.8 mm ID X 30.0 cm L	G6000 PW <sub>XL</sub> -CP	13 µm
	21876	6.0 mm ID X 4.0 cm L	PW <sub>XL</sub> -CP Guardcolumn	12 µm

This sheet contains the recommended operating conditions and the specifications for TSK-GEL XL PW<sub>XL</sub>-CP columns and guard columns. Installation instructions and column care information are described in a separate Instruction Manual.

#### A. OPERATING CONDITIONS

1. Shipping Solvent: Water
2. Max. Flow Rate: 1.0 mL/min  
**Note:** When a buffer with high viscosity is used, the maximum flow rate may have to be reduced so as not to exceed the maximum pressure drop. When changing solvents, use a flow rate equal to 25% of the maximum flow rate.
3. Max. Pressure:
 

5.5 MPa	=	800	psi	P/N 21873
2.5 MPa	=	350	psi	P/N 21874
2.0 MPa	=	300	psi	P/N 21875
4. pH Range: 2.0 - 12.0
5. Salt Conc.: 0.5 Molar
6. Organic Conc.: 20% It is possible to use up to 50% organic when the solvent change is made very gradually using a shallow gradient at low flow rate.
7. Temperature: 10 - 60°C. Reduce flow rate when operating below 10°C.
8. Cleaning Solvents:
  - (1) High salt concentration buffer (0.5 - 1.0 M), or
  - (2) pH 2 - 3 or pH 9 - 12 buffer, or
  - (3) Buffer with acetonitrile or methanol, or
  - (4) Buffer with urea or SDS**Note:** Choose a cleaning solvent based on sample properties, e.g. use (1) to remove basic polymers, (3) to remove hydrophobic proteins etc.
9. Storage: Store the column in a 0.05% NaN<sub>3</sub> solution or 20% ethanol in DI water when it will not be used the next day. For overnight storage flush the column at low flow rate with the mobile phase. Prevent air from entering the column!
10. Column Protection: The use of a guard column (TSKgel PW<sub>XL</sub> Guard Column P/N 21876) is recommended to prolong the life of the analytical column. Guard column life depends greatly on sample cleanliness. As a general rule, guard columns should be replaced after every 30-40 sample injections, when the peaks become excessively wide, or when the peaks show splitting.

#### B. SPECIFICATIONS

The performance of TSK-GEL PW<sub>XL</sub> columns is tested under the conditions described in the Data Sheet. All columns have passed the following quality control specifications:

Number of Theoretical Plates (N):	≥	16,000	P/N 21873
	≥	10,000	P/N 21874
	≥	7,000	P/N 21875
Asymmetry Factor (AF):		0.7 - 1.6	