



Im Leuschnerpark 4, 64347 Griesheim, Germany  
 Tel: +49 6155-7043700 Fax: +49 6155-8357900  
 E-Mail: info.tbg@tosoh.com  
 Web: www.tosohbioscience.de

3604 Horizon Drive, Suite 100, King of Prussia, PA 19406, USA  
 Tel: +1 800-366-4875 Fax: +1 610-272-3028  
 E-Mail: info.tbl@tosoh.com  
 Web: www.tosohbioscience.com

## OPERATING CONDITIONS and SPECIFICATIONS

### TSKgel® SP-5PW Products

<b>Analytical/Semi-Prep:</b>	0018758	2.0 mm ID x 7.5 cm L		<b>10 µm</b>
	0007161	7.5 mm ID x 7.5 cm L		<b>10 µm</b>
	0007211		Guardgel Kit for PN 07161	<b>20 µm</b>
	0007575	21.5 mm ID x 15.0 cm L		<b>13 µm</b>
	0016093		Prep Guardgel Kit for PN 07575	<b>20 µm</b>
	0013062	5.0 mm ID x 5.0 cm L	Glass	<b>10 µm</b>
	0008803	8.0 mm ID x 7.5 cm L	Glass	<b>10 µm</b>
	0008807		Guardgel Kit Glass for PN 13062 & 08803	<b>20 µm</b>
	0014017	20.0 mm ID x 15.0 cm L	Glass	<b>13 µm</b>

**Small Ion Capacity:** >0.1meq/mL      **Counter Ion:** Na<sup>+</sup>

This sheet contains the recommended operating conditions and the specifications for **TSKgel** SP-5PW columns and guard columns. Installation instructions and column care information are described in a separate Instruction Manual.

#### A. OPERATING CONDITIONS

- Shipping Solvent: Distilled and Deionized Water
- Max.Flow Rate:
 

0.12 mL/min	2.0 mm ID x 7.5 cm L
1.0 mL/min	5.0 mm ID x 5.0 cm L, Glass
1.2 mL/min	7.5 mm ID x 7.5 cm L and 8.0 mm ID x 7.5 cm L Glass
8.0 mL/min	21.5 mm ID x 15.0 cm L and 20.0 mm ID x 15.0 cm L, Glass

**NOTE:** When changing solvents, use a flow rate equal to 50% of the maximum flow rate (20% for prep columns).

- Standard Flow Rate:
 

0.05 - 0.10 mL/min	2.0 mm ID x 7.5 cm L
0.50 - 0.80 mL/min	5.0 mm ID x 5.0 cm L, Glass
0.50 - 1.00 mL/min	7.5 mm ID x 7.5 cm L and 8.0 mm ID x 7.5 cm L Glass
4.00 - 6.00 mL/min	21.5 mm ID x 15.0 cm L and 20.0 mm ID x 15.0 cm L, Glass
- Max. Pressure:
 

1.0 MPa	8.0 mm ID x 7.5 cm L and 2.0 mm ID x 7.5 cm L
1.5 MPa	7.5 mm ID, 5.0 mm ID Glass and 20.0 mm ID Glass
2.5 MPa	21.5 mm ID x 15.0 cm L

- pH Range: 2.0 - 12.0
- Salt Conc.: < 1.0 Molar
- Organic Conc.: ≤ 20%

- Temperature: 10 - 45 °C      Reduce flow rate when operating below 10 °C.

- Cleaning Solvents:
  - (1) 0.1 - 0.2 M NaOH, and, if not successful,
  - (2) 20 - 40% acetic acid aq., or
  - (3) Aqueous buffer in organic solvent, or, if nothing else was successful,
  - (4) Urea or nonionic surfactant in buffer.

**NOTE:** Clean the column regularly by injecting up to one column volume of 0.1 - 0.2 N NaOH in 250µl increments. Use proportionally larger volumes for (semi-) prep columns.

- Storage: Store the column in DI water when it will not be used the next day or 20% ethanol or methanol in DI water for longer term storage. For overnight storage flush the column with the low salt concentration mobile phase at low flow rate. Prevent air from entering the column!

11. Column Protection:

The use of guard columns is recommended to prolong the life of the analytical column. Guard column life depends greatly on sample cleanliness. As a general rule, guard columns should be replaced when the peaks become excessively wide, or when the peaks show splitting.

**B. SPECIFICATIONS**

The performance of **TSKgel** SP-5PW columns is tested under the conditions described in the Data Sheet. All columns have passed the following quality control specifications

Number of Theoretical Plates (N):	$\geq$	700	5.0 mm ID Glass
	$\geq$	1,300	7.5 mm ID, 8.0 mm ID Glass and 2.0 mm ID
	$\geq$	3,000	21.5 mm ID and 20.0 mm ID Glass
Asymmetry Factor (AF):		0.8 - 1.6	for all analytical/semi-prep columns; no spec for prep columns