Im Leuschnerpark 4, 64347 Griesheim, Germany Tel: +49 6155-7043700 Fax: +49 6155-8357900

E-Mail: info.tbg@tosoh.com Web: www.tosohbioscience.de 3604 Horizon Drive, Suite 100, King of Prussia, PA 19406, USA

Tel: +1 800-366-4875 Fax: +1 610-272-3028

E-Mail: info.tbl@tosoh.com Web: www.tosohbioscience.com

OPERATING CONDITIONS and SPECIFICATIONS

TSKgel® ODS-80TM Products

Part Numbers: 16651 4.6 mm ID x 7.5 cm L 5 um 4.6 mm ID x 15.0 cm L 08148 5 µm 08149 4.6 mm ID x 25.0 cm L 5 µm 14002 21.5 mm ID x 30.0 cm L 10 µm Guard Filter Holder 18206 18207 Guard Filter (pk3) 14098 21.5 mm ID x 7.5 cm L **Guard Column**

This sheet contains the recommended operating conditions and the specifications for **TSKgel** ODS-80 TM columns and guard columns Installation instructions and column care information are described in a separate Instruction Manual.

A. OPERATING CONDITIONS

Shipping Solvent: 70% Methanol - 30% Water

2. Max.Flow Rate: 1.2 mL/min 4.6 mm ID

12.0 mL/min 21.5 mm ID

NOTE: When a buffer with high viscosity is used, the maximum flow rate may have to be

reduced so as not to exceed the maximum pressure drop. When changing solvents,

use a flow rate equal to 25% of the maximum flow rate.

s. Standard Flow Rate: 0.8 - 1.0 mL/min 4.6 mm ID

4.0 - 6.0 mL/min 21.5 mm ID

4. Max. Pressure: 10 MPa 7.5 cm L column

20 MPa
 30 MPa
 15.0 cm L column
 25.0 cm L column

6 MPa 21.5 mm ID

5. pH Range: 2.0 - 7.5

7. Organic Conc.: 0 - 100%

8. Temperature: 0 - 45°C Reduce flow rate when operating below 10°C

9. Cleaning Solvents: (1) Acetonitrile or Methanol, or

(2) gradient from 10% to 100% acetonitrile in 0.05% trifluoroacetic acid

Storage: Store the column in acetonitrile or methanol when it will not be used the next day.

For overnight storage flush the column with the mobile phase at 0.2 mL/min.

1. Column Protection: The use of guard columns is recommended to prolong the life of the analytical

The use of guard columns is recommended to prolong the life of the analytical column. Guard column life depends greatly on sample cleanliness. As a general rule, guard columns should be replaced after every 30 - 40 sample injections or when the peaks become excessively wide or when the peaks show splitting.

B. SPECIFICATIONS

The performance of TSKgel ODS-80 TM columns is tested under the conditions described in the

Data Sheet. All columns have passed the following quality control specifications

Number of Theoretical Plates (N): ≥ 4,500 7.5 cm L column

Asymmetry Factor (AF): 0.7 - 1.8 4.6 mm ID and 6.0 mm ID

Note our technical hotline tel +49 6155 70437-36 and e-mail, techsupport.tbg@tosoh.com

Page 1 of 1 DS 1043 / July 15 / AX