



Im Leuschnerpark 4, 64347 Griesheim, Germany
Tel: +49 6155-7043700 Fax: +49 6155-8357900
E-Mail: info.tbg@tosoh.com
Web: www.tosohbioscience.de

3604 Horizon Drive, Suite 100, King of Prussia, PA 19406, USA
Tel: +1 800-366-4875 Fax: +1 610-272-3028
E-Mail: info.tbl@tosoh.com
Web: www.tosohbioscience.com

OPERATING CONDITIONS and SPECIFICATIONS

TSKgel® CN-80TS Products

Part Numbers:	0017348	4.6 mm ID x 15.0 cm L	5 µm
	0017349	4.6 mm ID x 25.0 cm L	5 µm

This sheet contains the recommended operating conditions and the specifications for **TSKgel** CN-80TS columns and Guardcolumns. Installation instructions and column care information are described in a separate Instruction Manual.

A. OPERATING CONDITIONS

1. Shipping Solvent: 60% Acetonitrile - 40% Water

2. Max.Flow Rate: 1.2 mL/min

NOTE: When a buffer with high viscosity is used, the maximum flow rate may have to be reduced so as not to exceed the maximum pressure drop. When changing solvents, use a flow rate equal to 25% of the maximum flow rate.

3. Standard Flow Rate: 0.8 – 1.0 mL/min 4.6 mm ID, 6.0 mm ID

4. Max. Pressure: 20 MPa 4.6 mm ID x 15.0 cm L
30 MPa 4.6 mm ID x 25.0 cm L

5. pH Range: 2.0 - 7.5

6. Salt Conc.: No limitation

7. Organic Conc.: No limitation

8. Temperature: 10 - 50°C Reduce flow rate when operating below 10°C.

9. Cleaning Solvents: (1) High conc. solvent containing organic modifiers*
(2) Mixture of organic acids and high conc. organic modifiers*

NOTE: *Acetonitrile and methanol are recommended as a modifier.

10. Storage: Store the column in the shipping solvent if it will not be used within three days. Prevent air from entering the column. For overnight storage flush the column with mobile phase at 0.2 mL/min.

11. Column Protection: An on-line filter (0.2-0.5 µm) equipped between pump and injection valve is useful. Instead of a guardcolumn, which is not available for the **TSKgel** CN-80TS, an Pre-column Filter with 0.5 µm frits (P/N 0003410, P/N 0003411 for replacement frits) is recommended.

B. SPECIFICATIONS

The performance of **TSKgel** CN-80TS columns is tested under the conditions described in the Data Sheet. All columns have passed the following quality control specifications

Number of Theoretical Plates (N):	≥ 11,000	15 cm columns
	≥ 18,000	25 cm columns

Asymmetry Factor (AF): 0.7 - 1.8