TOSOH BIOSCIENCE

Im Leuschnerpark 4, 64347 Griesheim, Germany
Tel: +49 6155-7043700 Fax: +49 6155-8357900
E-Mail: info.tbg@tosoh.com
Web: www.tosohbioscience.de

3604 Horizon Drive, Suite 100, King of Prussia, PA 19406, USA
Tel: +1 800-366-4875 Fax: +1 610-272-3028
E-Mail: info.tbl@tosoh.com
Web: www.tosohbioscience.com

OPERATING CONDITIONS and SPECIFICATIONS

TSKgel ® SWXL Guardcolumn Products

Part Numbers:	0008543	6.0 mm ID x 4.0 cm L	Guardcolumn SWXL	**7 µm**
	0018008	6.0 mm ID x 4.0 cm L	Guardcolumn SWXL PEEK	**7 µm**

This sheet contains the recommended operating conditions and the specifications for **TSKgel** SWXL guardcolumns.
Installation instructions and column care information are described in a separate Instruction Manual.

A. OPERATING CONDITIONS

1.	Shipping Solvent:	0.05% NaN_3 and 0.1 M Na_2SO_4 in 0.1 M phosphate buffer, pH 6.7	
2.	Max.Flow Rate:	1.2	mL/min
	NOTE:		When a buffer with high viscosity is used, the maximum flow rate may have to be reduced so as not to exceed the maximum pressure drop. When changing solvents, use a flow rate equal to 25% of the maximum flow rate.
3.	Standard Flow Rate:	0.5 - 1.0	mL/min
4.	Max. Pressure:	7.0	MPa
5.	pH Range:	2.5 - 7.5	
6.	Salt Conc.:	≤ 0.5 Molar	
7.	Organic Conc.:	0 - 100%	for aqueous soluble organic solvents. Make gradual solvent changes using a shallow
8.	Temperature:	10 - 30°C	Reduce flow rate when operating below 10°C.
9.	Cleaning Solvents:		(1) conc. salt solution at low pH, e.g. 0.5 M Na_2SO_4, pH 2.7 (2) methanol or acetonitrile in low conc. aqueous buffer, or, if nothing else is successful, (3) buffered solution of SDS, urea or guanidine
	NOTE:		Choose a cleaning solvent based on sample properties, e.g. use (1) to remove basic proteins, and (2) to remove hydrophobic proteins. Chaotrophic agents can solvate strongly adsorbed proteins, e.g. via hydrogen bonding.
10.	Storage:		Store the column in mobile phase containing 0.05% NaN_3 when it will not be used the next day. For overnight storage flush the column with mobile phase at low flow rate. Prevent air from entering the column!
11.	Column Protection:		The use of guard columns is recommended to prolong the life of the analytical column. Guard column life depends greatly on sample cleanliness. As a general rule, guard columns should be replaced after every 30-40 sample injections, when the peaks become excessively wide, or when the peaks show splitting.
12.	Top-Off:		Occasionally, due to accident, sample, mobile phase or operational variables, a depression can develop at the guard column inlet. Use **TSKgel** Top-Off SWXL (P/N 08544) for filling in such voids.

Note our technical hotline tel +49 6155 70437-36
and e-mail, techsupport.tbg@tosoh.com

DS 1141 / July 15 / AX