



Im Leuschnerpark 4, 64347 Griesheim, Germany  
Tel: +49 6155-7043700 Fax: +49 6155-8357900  
E-Mail: info.tbg@tosoh.com  
Web: www.tosohbioscience.de

3604 Horizon Drive, Suite 100, King of Prussia, PA 19406, USA  
Tel: +1 800-366-4875 Fax: +1 610-272-3028  
E-Mail: info.tbl@tosoh.com  
Web: www.tosohbioscience.com

## OPERATING CONDITIONS and SPECIFICATIONS

### TSKgel® G5000PWXL Products

**Part Numbers:** 0008023 7.8mm ID x 30.0 cm L 10 µm

This sheet contains the recommended operating conditions and the specifications for **TSKgel** G5000PWXL columns. Installation instructions and column care information are described in a separate Instruction Manual.

#### A. OPERATING CONDITIONS

- |                        |   |
|------------------------|---|
| 1. Shipping Solvent:   | Water   |
| 2. Max. Flow Rate:     | 1.0 mL/min  |
| <b>NOTE:</b>           | When a buffer with high viscosity is used, the maximum flow rate may have to be reduced so as not to exceed the maximum pressure drop. When changing solvents, use a flow rate equal to 25% of the maximum flow rate.   |
| 3. Standard Flow Rate: | 0.3 - 0.6 mL/min  |
| 4. Max. Pressure:      | 2 MPa   |
| 5. pH Range:           | 2.0 - 12.0  |
| 6. Salt Conc.:         | ≤ 0.5 M   |
| 7. Organic Conc.:      | ≤ 20% It is possible to use up to 50% organic when the solvent change is made very gradually using a shallow gradient at low flow rate.   |
| 8. Temperature:        | 10 - 80°C Reduce flow rate when operating below 10°C.   |
| 9. Cleaning Solvents:  | (1) High salt concentration buffer (0.5 - 1.0 M), or<br>(2) pH 2 - 3 or pH 9 - 12 buffer, or<br>(3) Buffer with acetonitrile or methanol, or if not successful<br>(4) Buffer with urea or SDS   |
| <b>NOTE:</b>           | Choose a cleaning solvent based on sample properties, e.g. use (1) to remove basic polymers, (3) to remove hydrophobic proteins etc.  |
| 10. Storage:           | Store the column in a 0.05% NaN <sub>3</sub> solution or 20% ethanol in DI water when it will not be used the next day. For overnight storage flush the column at low flow rate with the mobile phase. Prevent air from entering the column!  |
| 11. Column Protection: | The use of a guard column ( <b>TSKgel</b> PWXL Guard Column P/N 0008033) is recommended to prolong the life of the analytical column. Guard column life depends greatly on sample cleanliness. As a general rule, guard columns should be replaced after every 30-40 sample injections, when the peaks become excessively wide, or when the peaks show splitting. |
| 12. Top-Off:           | Occasionally, due to accident, sample, mobile phase or operational variables, a depression can develop at the column or guard column inlet.<br>Use <b>TSKgel</b> Top-Off PWXL (P/N 0008035) for filling in such voids.  |

#### B. SPECIFICATIONS

The performance of **TSKgel** G5000PWXL columns is tested under the conditions described in the Data Sheet. All columns have passed the following quality control specifications

Number of Theoretical Plates (N):	≥ 10,000
Asymmetry Factor (AF):	0.7 - 1.6