



Im Leuschnerpark 4, 64347 Griesheim, Germany
 Tel: +49 6155-7043700 Fax: +49 6155-8357900
 E-Mail: info.tbg@tosoh.com
 Web: www.tosohbioscience.de

3604 Horizon Drive, Suite 100, King of Prussia, PA 19406, USA
 Tel: +1 800-366-4875 Fax: +1 610-272-3028
 E-Mail: info.tbl@tosoh.com
 Web: www.tosohbioscience.com

OPERATING CONDITIONS and SPECIFICATIONS

TSKgel® ODS-100Z

Part Numbers:	0022726	2.0 mm ID x 1.0 cm L	3 µm	3 pieces /pkg Guard Cartridges Holder recommended
	0022727	2.0 mm ID x 2.0 cm L	3 µm	0022748 2.0 mm ID x 1.0 cm L 5 µm
	0022728	2.0 mm ID x 3.5 cm L	3 µm	0022749 2.0 mm ID x 2.0 cm L 5 µm
	0022729	2.0 mm ID x 5.0 cm L	3 µm	0022750 2.0 mm ID x 3.5 cm L 5 µm
	0022730	2.0 mm ID x 7.5 cm L	3 µm	0021460 2.0 mm ID x 5.0 cm L 5 µm
	0022731	2.0 mm ID x 10.0 cm L	3 µm	0022751 2.0 mm ID x 7.5 cm L 5 µm
	0022732	2.0 mm ID x 15.0 cm L	3 µm	0022752 2.0 mm ID x 10.0 cm L 5 µm
	0022733	2.0 mm ID x 25.0 cm L	3 µm	0021459 2.0 mm ID x 15.0 cm L 5 µm
				0022753 2.0 mm ID x 25.0 cm L 5 µm
	0022734	3.0 mm ID x 2.0 cm L	3 µm	0022754 3.0 mm ID x 2.0 cm L 5 µm
	0022735	3.0 mm ID x 3.5 cm L	3 µm	0022755 3.0 mm ID x 3.5 cm L 5 µm
	0022736	3.0 mm ID x 5.0 cm L	3 µm	0022756 3.0 mm ID x 5.0 cm L 5 µm
	0022737	3.0 mm ID x 7.5 cm L	3 µm	0022757 3.0 mm ID x 7.5 cm L 5 µm
	0022738	3.0 mm ID x 10.0 cm L	3 µm	0022758 3.0 mm ID x 10.0 cm L 5 µm
	0022739	3.0 mm ID x 15.0 cm L	3 µm	0022759 3.0 mm ID x 15.0 cm L 5 µm
	0022740	3.0 mm ID x 25.0 cm L	3 µm	0022760 3.0 mm ID x 25.0 cm L 5 µm
	0022741	4.6 mm ID x 2.0 cm L	3 µm	0022761 4.6 mm ID x 2.0 cm L 5 µm
	0022742	4.6 mm ID x 3.5 cm L	3 µm	0022762 4.6 mm ID x 3.5 cm L 5 µm
	0022743	4.6 mm ID x 5.0 cm L	3 µm	0022763 4.6 mm ID x 5.0 cm L 5 µm
	0022744	4.6 mm ID x 7.5 cm L	3 µm	0022764 4.6 mm ID x 7.5 cm L 5 µm
	0022745	4.6 mm ID x 10.0 cm L	3 µm	0022765 4.6 mm ID x 10.0 cm L 5 µm
	0022746	4.6 mm ID x 15.0 cm L	3 µm	0022461 4.6 mm ID x 15.0 cm L 5 µm
	0022747	4.6 mm ID x 25.0 cm L	3 µm	0022462 4.6 mm ID x 25.0 cm L 5 µm
	0021996	2.0 mm ID x 1.0 cm L	3 µm	pkg. of 3 Guard Cartridges for all 3 µm ODS-100Z 2.0 mm ID columns
	0021995	2.0 mm ID x 1.0 cm L	5 µm	pkg. of 3 Guard Cartridges for all 5 µm ODS-100Z 2.0 mm ID columns
	0021454	3.2 mm ID x 1.5 cm L	5 µm	pkg. of 3 Guard Cartridges for all ODS-100V 4.6 mm ID columns
	0019308	TSKgel Guard Cartridge Holder for 2.0 mm ID Cartridges		
	0019018	TSKgel Guard Cartridge Holder for 3.2 mm ID Cartridges		

This sheet contains the recommended operating conditions and the specifications for **TSKgel** ODS-100V columns. Installation instructions and column care information are described in a separate Instruction Manual.

A. OPERATING CONDITIONS

1. Shipping Solvent: Acetonitrile

Note: When a buffer with high viscosity is used, the flow rate may have to be reduced so as not to exceed the maximum pressure drop. When changing solvents, use a flow rate equal to 25% of the maximum flow rate.

2. Max. Pressure:		Part Numbers
	9 MPa	P/N 0022749; P/N 0022750; P/N 0022754; P/N 0022755; P/N 0022761; P/N 0022762;
	12 MPa	P/N 0022727; P/N 0022734; P/N 0022735; P/N 0022741; P/N 0022742; P/N 0022756; P/N 0022763;
	15 MPa	P/N 0022728; P/N 0022729; P/N 0022736; P/N 0022743;
	18 MPa	P/N 0022757; P/N 0022758; P/N 0022759; P/N 0022760; P/N 0022751; P/N 0022752; P/N 0021459; P/N 0022753; P/N 0022760; P/N 0022764; P/N 0022765; P/N 0022761;
	21 MPa	P/N 0022730; P/N 0022737; P/N 0022744; P/N 0021462;
	24 MPa	P/N 0022731; P/N 0022732; P/N 0022738; P/N 0022739; P/N 0022745; P/N 0022746;
	28 MPa	P/N 0022748;
	30 MPa	P/N 0022726; P/N 0022733; P/N 0022740; P/N 0022747;

Note our technical hotline tel +49 6155 70437-36 and e-mail, techsupport.tbg@tosoh.com

3. pH Range: 2.0 - 7.5
4. Organic Conc.:
 5 μ m: 0 – 100% polar and non polar switch to 100% acetonitrile before switching to 100% non polar solvents.
 3 μ m: 0 – 100% polar and non polar switch to 95% acetonitrile before switching to 100% non polar solvents.
5. Temperature:
 10°C - 80°C 5 μ m columns
 10°C - 50°C 3 μ m columns
6. Cleaning Solvents
 (1) High conc. solvent containing organic modifiers*
 (2) Mixture of organic acids and high conc. organic modifiers*
 *Acetonitrile and methanol are recommended as a modifier.
7. Storage: Store the column in the shipping solvent when it will not be used the next day. For overnight storage flush the column with the mobile phase at 0.2 mL/min (4.6 mm ID) or 0.05 mL/min (2.0 mm ID) respectively.
8. Column Protection: An on-line filter (0.2-0.5 μ m) equipped between pump and injection valve is recommended. Guard columns prevent the column from a contamination of strongly adsorbed solutes. As a general rule, guard columns should be replaced after 30-40 sample injections or when peaks become excessively wide.

B. SPECIFICATIONS

The performance of **TSKgel** ODS-100V column is tested under the conditions described in the data sheet. All columns have passed the following quality control specifications:

No. of Theoretical Plates (N):	Part Numbers	TP (N)	Part Numbers
\geq 300	P/N 0022748	\geq 9,000	P/N 0022737; P/N 0022765;
\geq 500	P/N 0022726;	\geq 9,750	P/N 0022744;
\geq 1,000	P/N 0022749;	\geq 11,000	P/N 0021459;
\geq 1,200	P/N 0022754;	\geq 11,500	P/N 0022731;
\geq 1,500	P/N 0022727; P/N 0022761;	\geq 12,000	P/N 0022738;
\geq 2,000	P/N 0022734;	\geq 13,000	P/N 0022759;
\geq 2,500	P/N 0022741; P/N 0022750;	\geq 13,500	P/N 0022745;
\geq 3,000	P/N 0021460; P/N 0022755; P/N 0022762;	\geq 14,000	P/N 0021461;
\geq 4,000	P/N 0022728; P/N 0022735; P/N 0022756;	\geq 17,500	P/N 0021810; P/N 0022732;
\geq 4,500	P/N 0022742; P/N 0022763;	\geq 18,000	P/N 0022739; P/N 0022753;
\geq 5,500	P/N 0022751;	\geq 19,500	P/N 0022746;
\geq 5,700	P/N 0022729;	\geq 21,000	P/N 0022760;
\geq 6,000	P/N 0022736; P/N 0022757;	\geq 23,000	P/N 0021462;
\geq 6,500	P/N 0022743;	\geq 28,000	P/N 0022733;
\geq 7,000	P/N 0022752; P/N 0022764;	\geq 29,000	P/N 0022740;
\geq 8,500	P/N 0022758;	\geq 30,000	P/N 0022747;
\geq 8,600	P/N 0022730;		

Asymmetry Factor (AF):	Part Numbers	(AF):	Part Numbers		
0.90 - 1.20	P/N 0022730; P/N 0022731; P/N 0022732; P/N 0022733; P/N 0022735; P/N 0022737; P/N 0022738; P/N 0022739; P/N 0022740; P/N 0022742; P/N 0022744; P/N 0022745; P/N 0022746; P/N 0022747;	3 μm	0.90 - 1.15	P/N 0021461; P/N 0021462;	5 μm
			0.90 - 1.20	P/N 0022750; P/N 0021460; P/N 0022751; P/N 0022751; P/N 0021459; P/N 0022753; P/N 0022755; P/N 0022756; P/N 0022757; P/N 0022758; P/N 0022759; P/N 0022760; P/N 0022762; P/N 0022763; P/N 0022764; P/N 0022765;	5 μm
0.90 - 1.30	P/N 0022729; P/N 0022736; P/N 0022743;	3 μm			
0.90 - 1.40	P/N 0022726;	3 μm			
0.90 - 1.60	P/N 0022727; P/N 0022733; P/N 0022741	3 μm	0.90 - 1.60	P/N 0022749; P/N 0022754; P/N 0022761;	5 μm
0.90 - 2.20	P/N 0022728;	3 μm	0.90 - 2.20	P/N 0022748;	5 μm