



Im Leuschnerpark 4, 64347 Griesheim, Germany
 Tel: +49 6155-7043700 Fax: +49 6155-8357900
 E-Mail: info.tbq@tosoh.com
 Web: www.tosohbioscience.de

3604 Horizon Drive, Suite 100, King of Prussia, PA 19406, USA
 Tel: +1 800-366-4875 Fax: +1 610-272-3028
 E-Mail: info.tbl@tosoh.com
 Web: www.tosohbioscience.com

OPERATING CONDITIONS and SPECIFICATIONS

TSKgel® STAT Series Anion Exchange Products

Part Numbers:	0021960	3.0 mm ID X 3.5 cm L	TSKgel Q-STAT	10 µm
	0021961	4.6 mm ID X 10.0 cm L	TSKgel Q-STAT	7 µm
	0021962	4.6 mm ID X 10.0 cm L	TSKgel DNA-STAT	5 µm

This sheet contains the recommended operating conditions and the specifications for **TSKgel** STAT Series Anion Exchange columns. Installation instructions and column care information are described in a separate Instruction Manual.

A. OPERATING CONDITIONS

1. Shipping Solvent: Ion-Exchanged Water
2. Standard Flow Rate:

1.0 - 2.0	mL/min	P/N 0021960
0.5 - 1.4	mL/min	P/N 0021961
0.3 - 0.6	mL/min	P/N 0021962
- NOTE:** When a buffer with high viscosity is used, the maximum flow rate may have to be reduced so as not to exceed the maximum pressure drop. When changing solvents, use a flow rate equal to 25% of the maximum flow rate.
3. Max. Pressure:

10	MPa	P/N 0021960
10	MPa	P/N 0021961
15	MPa	P/N 0021962
4. pH Range: 3.0 - 10.0
5. Salt Conc.: No Limitation
6. Organic Conc.: <50% When solvent in column is replaced by distilled or ion-exchanged water, feed the solvent slowly, at flow rates <0.5 mL/min (Q-STAT) and <0.25 mL/min (DNA-STAT).
7. Temperature: 10 - 60°C
8. Cleaning Solvents: Adsorbed materials can be stripped from the column by repeated injection with one of the following cleaning solutions:
 - (1) 0.1 M NaOH, or
 - (2) 20 ~ 40% Acetic acid, or
 - (3) Solution containing aqueous organic solvent such as methanol or acetonitrile, or
 - (4) Solution containing a solubilizer such as urea and non-ionic surfactants
9. Storage: Short term storage: keep the column filled with low ionic strength eluent.
For long term storage, replace the solvent in the column with distilled or ion-exchanged water, at flow rates < 0.5 ml/min (Q-STAT) and < 0.25 ml/min (DNA-STAT).

B. SPECIFICATIONS

The performance of **TSKgel** Q-STAT and **TSKgel** DNA-STAT columns is tested under the conditions described in the Data Sheet. All columns have passed the following quality control specifications

Number of Theoretical Plates (N):	≥	200	P/N 0021960
	≥	4,000	P/N 0021961
	≥	4,000	P/N 0021962
Asymmetry Factor (AF):		0.8 - 1.8	P/N 0021960
		1.0 - 2.0	P/N 0021961
		1.0 - 2.0	P/N 0021962